
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005888
11:13 04/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:34
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 225kPa 4:08
136.4 C 225kPa 4:38
136.4 C 224kPa 5:08
136.0 C 223kPa 5:38
136.2 C 223kPa 6:08
136.4 C 223kPa 6:38
136.4 C 220kPa 7:08
136.4 C 220kPa 7:08
VENTING 7:08
AIR DRYING 8:15
CYCLE COMPLETE 11:02

Digital Signature #
22A498FAF1FFCED8
