
Midmark M11 - v1.0.5
Total Cycles: 143

Name: RITTER2
Sterilizer ID: V2313228
06/25/2021 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 150.5 F 1.7
4:00 157.3 F 4.7
6:00 178.2 F 8.2
8:00 199.0 F 12.4
10:00 214.5 F 5.6
12:00 223.9 F 9.9
14:00 231.8 F 9.6
16:00 238.4 F 11.9
18:00 239.0 F 10.6
20:00 250.2 F 16.3
22:00 260.7 F 22.2
24:00 270.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.5
1:00 272.7 F 30.3
2:00 272.4 F 30.0
3:00 272.1 F 29.5
4:00 271.7 F 29.7

Min 271.5 F 29.4
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 230.1 F 0.0
10:00 219.2 F 0.0
15:00 207.9 F 0.0
20:00 198.3 F 0.0
25:00 188.5 F 0.0
30:00 179.7 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 25:03
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

06/25/2021 11:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____