
Midmark M11 - v1.0.5
Total Cycles: 650

Name: RITTER2
Sterilizer ID: V2313228
12/20/2021 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 180.3 F 1.5
4:00 185.2 F 5.1
6:00 199.5 F 9.9
8:00 223.5 F 6.9
10:00 230.9 F 9.5
12:00 234.5 F 9.9
14:00 241.4 F 12.2
16:00 240.6 F 11.1
18:00 253.2 F 17.5
20:00 264.2 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 271.9 F 29.0
2:00 272.2 F 29.2
3:00 272.2 F 29.5
4:00 271.8 F 29.7

Min 271.4 F 29.0
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 222.6 F 0.0
10:00 189.3 F 0.0
15:00 172.7 F 0.0
20:00 171.2 F 0.0
25:00 178.5 F 0.0
30:00 175.8 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:40
CYCLE COMPLETE

12/20/2021 10:06 AM
ACCEPTED: X
REJECTED:
Operator: U02