
Midmark M11 - v1.0.5
Total Cycles: 1328

Name: RITTER2
Sterilizer ID: V2313228
07/21/2022 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 158.1 F 0.0
4:00 165.6 F 1.3
6:00 190.4 F 6.0
8:00 210.2 F 10.8
10:00 220.7 F 6.2
12:00 232.7 F 10.8
14:00 232.1 F 7.9
16:00 242.1 F 11.7
18:00 245.0 F 13.4
20:00 257.4 F 20.2
22:00 268.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.3
1:00 271.6 F 29.8
2:00 272.4 F 29.7
3:00 271.2 F 29.2
4:00 271.9 F 30.1

Min 271.1 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 211.9 F 0.0
10:00 229.9 F 0.0
15:00 233.0 F 0.0
20:00 226.1 F 0.0
25:00 207.6 F 0.0
30:00 204.2 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 23:05
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

07/21/2022 09:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____