
Midmark M11 - v1.0.0
Total Cycles: 2356

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2022 06:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.7 F 0.0
2:00 77.5 F 0.0
4:00 84.9 F 0.0
6:00 119.4 F 0.1
8:00 154.8 F 2.8
10:00 180.9 F 6.6
12:00 192.1 F 10.5
14:00 224.1 F 6.0
16:00 227.0 F 6.7
18:00 236.6 F 10.6
20:00 235.4 F 8.1
22:00 243.4 F 12.3
24:00 240.4 F 10.5
26:00 253.5 F 17.0
28:00 264.6 F 23.6
30:00 271.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.9 F 28.5
2:00 272.3 F 28.8
3:00 272.6 F 29.1
4:00 271.9 F 29.0

Min 271.5 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 235.6 F 0.0
10:00 210.2 F 0.0
15:00 218.5 F 0.0
20:00 225.7 F 0.0
25:00 227.8 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 30:37
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:46
CYCLE COMPLETE

09/22/2022 08:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____