
Midmark M11 - v1.0.5
Total Cycles: 1079

Name: RITTER2
Sterilizer ID: V2313228
05/05/2022 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.2 F 0.0
2:00 126.3 F 0.0
4:00 150.2 F 2.2
6:00 181.3 F 6.3
8:00 199.1 F 11.2
10:00 220.3 F 6.1
12:00 234.7 F 11.2
14:00 231.8 F 8.2
16:00 240.6 F 11.4
18:00 244.7 F 13.0
20:00 256.8 F 19.5
22:00 268.3 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.6 F 29.4
2:00 271.6 F 29.3
3:00 271.7 F 29.2
4:00 272.0 F 29.3

Min 271.5 F 29.2
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 232.5 F 0.0
10:00 215.5 F 0.0
15:00 180.9 F 0.0
20:00 178.6 F 0.0
25:00 144.9 F 0.0
30:00 138.3 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:53
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

05/05/2022 02:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____