
Midmark M11 - v1.0.5
Total Cycles: 1430

Name: RITTER2
Sterilizer ID: V2313228
08/30/2022 07:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.4 F 0.0
2:00 159.6 F 0.8
4:00 178.3 F 3.4
6:00 198.5 F 8.0
8:00 215.0 F 13.1
10:00 228.0 F 8.6
12:00 236.9 F 10.9
14:00 241.4 F 12.2
16:00 240.7 F 11.2
18:00 253.8 F 18.0
20:00 265.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.6 F 29.8
2:00 271.6 F 29.0
3:00 272.2 F 30.0
4:00 272.3 F 29.5

Min 271.2 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 229.1 F 0.0
10:00 223.5 F 0.0
15:00 223.2 F 0.0
20:00 217.2 F 0.0
25:00 214.1 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

08/30/2022 08:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____