

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3907

Name: RITTER 1  
Sterilizer ID: V2178198  
10/16/2023 12:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.6 F 0.0  
2:00 192.0 F 3.8  
4:00 193.4 F 8.4  
6:00 219.4 F 7.2  
8:00 227.2 F 8.7  
10:00 238.4 F 10.9  
12:00 241.4 F 11.8  
14:00 239.7 F 10.3  
16:00 253.3 F 17.1  
18:00 265.0 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.0  
2:00 271.4 F 28.3  
3:00 272.6 F 29.1  
4:00 271.8 F 28.5  
  
Min 271.4 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 226.2 F 0.0  
10:00 237.9 F 0.0  
15:00 229.0 F 0.0  
20:00 231.5 F 0.0  
25:00 230.3 F 0.0  
30:00 219.6 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 19:49  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:02  
CYCLE COMPLETE

10/16/2023 01:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_