
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000996
12:55 16/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:01
STERILIZING 3:57
136.2 C 222kPa 3:57
136.2 C 222kPa 4:27
136.4 C 224kPa 4:57
136.2 C 220kPa 5:27
136.3 C 224kPa 5:57
136.4 C 222kPa 6:27
136.1 C 217kPa 6:57
136.3 C 225kPa 7:27
136.3 C 225kPa 7:27
VENTING 7:27
AIR DRYING 8:23
CYCLE COMPLETE 8:26

Digital Signature #
8DA75CFC5C050566
