

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3014

Name: RITTER2  
Sterilizer ID: V2313228  
10/16/2023 12:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.9 F 0.0  
2:00 191.7 F 3.8  
4:00 202.6 F 9.3  
6:00 220.9 F 7.4  
8:00 232.5 F 11.2  
10:00 234.5 F 9.2  
12:00 237.4 F 9.4  
14:00 249.5 F 15.6  
16:00 262.2 F 22.9  
18:00 272.7 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.6 F 29.8  
2:00 271.6 F 29.0  
3:00 272.6 F 30.0  
4:00 272.3 F 29.5  
  
Min 271.5 F 29.0  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 219.2 F 0.0  
10:00 218.0 F 0.0  
15:00 219.0 F 0.0  
20:00 219.9 F 0.0  
25:00 219.6 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 18:16  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:28  
CYCLE COMPLETE

10/16/2023 01:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_