
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002354
8:32 05/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:36
STERILIZING 3:46
136.3 C 222kPa 3:46
136.3 C 224kPa 4:17
136.0 C 220kPa 4:47
136.1 C 223kPa 5:17
136.1 C 221kPa 5:47
136.3 C 222kPa 6:17
136.4 C 223kPa 6:47
136.4 C 221kPa 7:17
136.4 C 221kPa 7:17
VENTING 7:17
AIR DRYING 8:06
CYCLE COMPLETE 38:06

Digital Signature #
DD4C9C3C82FE6A1B
