

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000827  
14:52 30/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:33  
PRESSURIZING 6:15  
STERILIZING 7:41  
136.0 C 220kPa 7:41  
136.3 C 222kPa 8:11  
135.9 C 219kPa 8:41  
136.4 C 223kPa 9:11  
135.9 C 219kPa 9:41  
136.4 C 223kPa 10:11  
136.4 C 220kPa 10:41  
136.4 C 221kPa 11:11  
136.4 C 221kPa 11:11  
VENTING 11:11  
AIR DRYING 12:19  
CYCLE COMPLETE 31:55

-----  
Digital Signature #  
8F8944C2146214A0  
-----