
Midmark M11 - v1.0.0
Total Cycles: 1970

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2022 09:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.8 F 0.0
2:00 131.7 F 1.2
4:00 140.7 F 3.9
6:00 190.1 F 7.7
8:00 211.5 F 12.2
10:00 210.8 F 5.2
12:00 222.5 F 9.6
14:00 232.2 F 9.6
16:00 235.5 F 10.1
18:00 235.2 F 8.2
20:00 246.5 F 13.7
22:00 258.2 F 20.0
24:00 268.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.7 F 29.4
2:00 271.9 F 28.7
3:00 272.1 F 29.1
4:00 272.6 F 29.2

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.5 F 0.0
5:00 202.9 F 0.0
10:00 217.8 F 0.0
15:00 201.6 F 0.0
20:00 221.2 F 0.0
25:00 236.7 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 25:15
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:03:14
CYCLE COMPLETE

05/30/2022 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____