

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4172

Name: RITTER2  
Sterilizer ID: V2313228  
12/02/2024 10:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.0 F 0.0  
2:00 72.1 F 0.0  
4:00 74.9 F 0.0  
6:00 97.5 F 0.0  
8:00 123.0 F 0.2  
10:00 159.6 F 0.3  
12:00 183.7 F 0.7  
14:00 198.5 F 2.4  
16:00 207.7 F 5.1  
18:00 220.6 F 7.9  
20:00 230.4 F 11.1  
22:00 228.6 F 7.9  
24:00 237.5 F 10.7  
26:00 236.3 F 8.8  
28:00 244.9 F 13.1  
30:00 242.6 F 11.8  
32:00 253.5 F 17.3  
34:00 263.1 F 22.9  
36:00 272.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.6  
1:00 271.6 F 29.0  
2:00 272.1 F 29.1  
3:00 271.3 F 28.4  
4:00 272.0 F 29.3  
  
Min 271.2 F 28.4  
Max 272.4 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 220.8 F 0.0  
10:00 209.9 F 0.0  
15:00 196.5 F 0.0  
20:00 198.9 F 0.0  
25:00 201.0 F 0.0  
30:00 206.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 36:25  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:15:01  
CYCLE COMPLETE

12/02/2024 11:37 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_