
Midmark M11 - v1.0.0
Total Cycles: 5073

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2024 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 182.8 F 0.3
4:00 185.0 F 2.8
6:00 197.4 F 7.5
8:00 213.8 F 12.3
10:00 227.6 F 7.2
12:00 237.4 F 12.5
14:00 238.1 F 10.2
16:00 236.0 F 8.4
18:00 250.3 F 15.4
20:00 262.1 F 22.2
22:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 29.1
2:00 271.9 F 28.6
3:00 271.1 F 28.4
4:00 272.4 F 29.2

Min 271.0 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 225.1 F 0.0
10:00 227.2 F 0.0
15:00 233.7 F 0.0
20:00 237.1 F 0.0
25:00 230.9 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

11/08/2024 12:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____