
Midmark M11 - v1.0.5
Total Cycles: 2240

Name: RITTER2
Sterilizer ID: V2313228
04/12/2023 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.2 F 0.0
2:00 201.0 F 4.7
4:00 216.9 F 11.2
6:00 228.8 F 8.1
8:00 238.3 F 10.8
10:00 244.4 F 13.4
12:00 245.0 F 13.1
14:00 258.8 F 20.5
16:00 270.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.6 F 29.4
2:00 271.4 F 29.1
3:00 272.6 F 29.4
4:00 271.4 F 28.8

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 228.5 F 0.0
10:00 239.1 F 0.0
15:00 228.6 F 0.0
20:00 226.5 F 0.0
25:00 229.5 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 2:53
HEATING: 16:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:56:49
CYCLE COMPLETE

04/12/2023 12:49 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____