
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002579
7:45 17/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:04
STERILIZING 3:10
136.4 C 223kPa 3:10
136.2 C 222kPa 3:40
136.4 C 224kPa 4:10
136.4 C 225kPa 4:40
136.0 C 221kPa 5:10
136.4 C 223kPa 5:40
136.4 C 224kPa 6:10
136.4 C 225kPa 6:40
136.4 C 225kPa 6:40
VENTING 6:40
AIR DRYING 7:32
CYCLE COMPLETE 37:32

Digital Signature #
9EF9F0F3DC03EDBC
