

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4097

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2024 01:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.1 F 0.0  
2:00 138.0 F 0.0  
4:00 163.5 F 1.7  
6:00 193.4 F 5.4  
8:00 226.1 F 9.3  
10:00 218.6 F 13.5  
12:00 229.8 F 7.2  
14:00 239.7 F 11.5  
16:00 234.4 F 8.5  
18:00 242.2 F 11.9  
20:00 245.2 F 13.0  
22:00 256.7 F 19.1  
24:00 267.2 F 25.6  
26:00 273.3 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.8 F 29.6  
2:00 271.5 F 29.3  
3:00 272.7 F 29.7  
4:00 271.5 F 29.1

Min 271.4 F 29.0  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.8 F 0.0  
5:00 229.2 F 0.0  
10:00 221.2 F 0.0  
15:00 223.9 F 0.0  
20:00 223.6 F 0.0  
25:00 230.3 F 0.0  
30:00 217.7 F 0.0

Drying Complete

FILLING: 0:57  
HEATING: 26:18  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:05:09  
CYCLE COMPLETE

11/07/2024 02:17 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_