
Midmark M11 - v1.0.0
Total Cycles: 4390

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2024 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 159.0 F 1.4
4:00 155.5 F 4.4
6:00 176.6 F 7.7
8:00 195.8 F 11.6
10:00 211.3 F 3.9
12:00 218.4 F 7.9
14:00 229.2 F 12.0
16:00 231.3 F 7.9
18:00 241.6 F 13.0
20:00 240.9 F 11.0
22:00 252.6 F 16.6
24:00 262.7 F 22.5
26:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.7 F 29.3
2:00 271.5 F 28.8
3:00 272.4 F 29.1
4:00 271.6 F 29.0

Min 271.5 F 28.7
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 230.9 F 0.0
10:00 228.9 F 0.0
15:00 235.3 F 0.0
20:00 223.3 F 0.0
25:00 220.9 F 0.0
30:00 221.0 F 0.0

Drying Complete

FILLING: 0:60
HEATING: 27:08
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:05:59
CYCLE COMPLETE

03/21/2024 11:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____