
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000799
8:26 04/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 3:26
STERILIZING 4:37
136.0 C 230kPa 4:37
136.3 C 233kPa 5:07
136.2 C 232kPa 5:37
136.1 C 230kPa 6:07
136.4 C 233kPa 6:37
136.3 C 232kPa 7:07
136.4 C 233kPa 7:37
136.4 C 233kPa 8:07
136.4 C 233kPa 8:07
VENTING 8:07
AIR DRYING 9:04
CYCLE COMPLETE 9:05

Digital Signature #
FE13103748D9BAF1
