
Midmark M11 - v1.0.5
Total Cycles: 274

Name: RITTER2
Sterilizer ID: V2313228
08/10/2021 06:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 74.9 F 0.0
4:00 81.9 F 0.6
6:00 116.1 F 2.8
8:00 149.5 F 5.5
10:00 173.3 F 8.5
12:00 189.2 F 12.4
14:00 206.3 F 4.6
16:00 217.5 F 8.0
18:00 228.3 F 11.8
20:00 230.6 F 7.8
22:00 240.5 F 12.8
24:00 239.7 F 11.0
26:00 251.7 F 16.9
28:00 262.0 F 22.8
30:00 271.9 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.1 F 29.0
2:00 272.7 F 29.5
3:00 271.9 F 29.7
4:00 271.3 F 28.9

Min 271.2 F 28.9
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.8 F 0.0
5:00 224.0 F 0.0
10:00 191.1 F 0.0
15:00 185.4 F 0.0
20:00 187.3 F 0.0
25:00 190.8 F 0.0
30:00 172.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

08/10/2021 07:35 AM
ACCEPTED: X
REJECTED:
Operator: U01