
Midmark M11 - v1.0.5
Total Cycles: 760

Name: RITTER2
Sterilizer ID: V2313228
02/07/2022 07:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 75.7 F 0.0
4:00 104.3 F 0.0
6:00 136.6 F 0.1
8:00 159.3 F 0.3
10:00 178.3 F 3.1
12:00 193.2 F 6.4
14:00 208.1 F 10.4
16:00 219.1 F 7.2
18:00 228.2 F 8.9
20:00 238.0 F 13.5
22:00 238.6 F 10.9
24:00 236.8 F 9.1
26:00 250.1 F 15.6
28:00 261.0 F 22.0
30:00 271.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.7
2:00 271.8 F 28.8
3:00 272.1 F 29.5
4:00 272.4 F 29.3

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 205.3 F 0.0
10:00 226.7 F 0.0
15:00 214.7 F 0.0
20:00 214.3 F 0.0
25:00 215.2 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:43
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:08:40
CYCLE COMPLETE

02/07/2022 08:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____