
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004008
8:50 14/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:06
STERILIZING 3:57
136.0 C 220kPa 3:57
136.2 C 222kPa 4:27
136.4 C 224kPa 4:57
136.4 C 223kPa 5:27
135.9 C 220kPa 5:57
136.4 C 223kPa 6:27
136.1 C 219kPa 6:57
136.3 C 221kPa 7:27
136.3 C 221kPa 7:27
VENTING 7:27
AIR DRYING 8:14
CYCLE COMPLETE 8:20

Digital Signature #
AC66E759573A973A
