

-----  
Midmark M11 - v1.0.5  
Total Cycles: 870

Name: RITTER2  
Sterilizer ID: V2313228  
03/10/2022 09:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.8 F 0.0  
2:00 178.8 F 3.3  
4:00 192.9 F 8.1  
6:00 211.9 F 12.2  
8:00 232.7 F 9.6  
10:00 232.0 F 8.9  
12:00 245.4 F 13.5  
14:00 244.5 F 12.6  
16:00 257.4 F 19.5  
18:00 268.8 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 271.8 F 29.2  
2:00 272.1 F 29.2  
3:00 272.7 F 29.4  
4:00 272.7 F 29.7  
  
Min 271.5 F 29.1  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
5:00 222.6 F 0.0  
10:00 224.7 F 0.0  
15:00 236.4 F 0.0  
20:00 226.7 F 0.0  
25:00 228.7 F 0.0  
30:00 226.3 F 0.0  
  
Drying Complete

FILLING: 2:13  
HEATING: 19:47  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:09  
CYCLE COMPLETE

03/10/2022 11:08 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02