
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002623
13:58 27/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:17
STERILIZING 3:18
136.0 C 222kPa 3:18
135.9 C 221kPa 3:48
136.3 C 225kPa 4:18
136.4 C 226kPa 4:48
136.1 C 223kPa 5:18
136.2 C 223kPa 5:48
136.1 C 223kPa 6:18
136.4 C 225kPa 6:48
136.4 C 225kPa 6:48
VENTING 6:49
AIR DRYING 7:37
CYCLE COMPLETE 20:47

Digital Signature #
AF6E90687BE8DA97
