
Midmark M11 - v1.0.0
Total Cycles: 788

Name: RITTER 1
Sterilizer ID: V2178198
04/15/2021 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.7 F 0.0
2:00 96.9 F 0.0
4:00 111.5 F 0.1
6:00 146.7 F 0.3
8:00 174.4 F 2.3
10:00 189.9 F 4.6
12:00 203.4 F 7.1
14:00 224.7 F 10.3
16:00 236.0 F 9.0
18:00 230.6 F 9.6
20:00 240.7 F 10.5
22:00 242.3 F 11.5
24:00 238.5 F 9.7
26:00 250.8 F 15.5
28:00 261.1 F 21.4
30:00 270.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 28.9
2:00 271.8 F 29.0
3:00 272.1 F 28.7
4:00 272.2 F 29.2

Min 271.5 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 231.1 F 0.0
10:00 220.6 F 0.0
15:00 190.7 F 0.0
20:00 179.4 F 0.0
25:00 197.0 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:10:19
CYCLE COMPLETE

04/15/2021 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____