
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000255
9:16 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:54
STERILIZING 5:25
136.4 C 233kPa 5:25
136.4 C 233kPa 5:55
136.3 C 232kPa 6:25
136.4 C 232kPa 6:55
136.2 C 231kPa 7:25
136.3 C 232kPa 7:55
136.2 C 232kPa 8:25
136.4 C 233kPa 8:55
136.4 C 233kPa 8:55
VENTING 8:55
AIR DRYING 9:58
CYCLE COMPLETE 24:58

Digital Signature #
ADF775395C28C206
