
Midmark M11 - v1.0.0
Total Cycles: 2315

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2022 06:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.4 F 0.0
2:00 83.2 F 0.0
4:00 114.8 F 0.0
6:00 156.2 F 0.0
8:00 177.6 F 0.1
10:00 196.6 F 0.6
12:00 209.1 F 1.6
14:00 220.5 F 5.1
16:00 228.7 F 8.0
18:00 237.0 F 11.5
20:00 234.6 F 7.2
22:00 241.3 F 12.8
24:00 239.2 F 10.7
26:00 237.2 F 9.0
28:00 251.0 F 15.8
30:00 262.4 F 22.4
32:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.3 F 28.7
2:00 272.5 F 29.1
3:00 271.7 F 28.3
4:00 272.8 F 29.1

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 189.8 F 0.0
10:00 249.4 F 0.0
15:00 213.8 F 0.0
20:00 215.9 F 0.0
25:00 221.4 F 0.0
30:00 224.4 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 32:20
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:10:25
CYCLE COMPLETE

09/13/2022 08:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____