

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004543  
13:29 17/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:38  
STERILIZING 3:34  
136.0 C 220kPa 3:34  
136.4 C 225kPa 4:04  
136.4 C 226kPa 4:34  
136.2 C 222kPa 5:04  
136.1 C 224kPa 5:34  
136.3 C 223kPa 6:04  
136.4 C 223kPa 6:34  
136.4 C 223kPa 7:04  
136.4 C 223kPa 7:04  
VENTING 7:04  
AIR DRYING 7:51  
CYCLE COMPLETE 8:03

-----  
Digital Signature #  
ABF995A2586807AA  
-----