
Midmark M11 - v1.0.0
Total Cycles: 4534

Name: RITTER 1
Sterilizer ID: V2178198
05/08/2024 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.1 F 0.0
2:00 145.5 F 1.8
4:00 151.7 F 4.7
6:00 172.4 F 8.1
8:00 214.1 F 12.5
10:00 214.8 F 5.1
12:00 225.9 F 9.2
14:00 237.9 F 12.0
16:00 236.9 F 10.4
18:00 235.7 F 8.5
20:00 247.6 F 14.1
22:00 258.3 F 20.0
24:00 268.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.1 F 28.8
2:00 272.7 F 29.2
3:00 272.3 F 29.4
4:00 271.5 F 28.7

Min 271.4 F 28.7
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 231.9 F 0.0
10:00 228.0 F 0.0
15:00 228.2 F 0.0
20:00 224.0 F 0.0
25:00 228.0 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:58
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:04:38
CYCLE COMPLETE

05/08/2024 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____