
Midmark M11 - v1.0.5
Total Cycles: 2818

Name: RITTER2
Sterilizer ID: V2313228
08/30/2023 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.5 F 0.0
2:00 167.4 F 0.2
4:00 194.7 F 4.4
6:00 211.7 F 9.1
8:00 226.0 F 9.9
10:00 233.7 F 10.5
12:00 233.5 F 8.3
14:00 240.1 F 10.6
16:00 247.7 F 14.5
18:00 260.2 F 21.5
20:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.4
2:00 272.1 F 29.5
3:00 272.2 F 29.2
4:00 272.3 F 29.8

Min 271.6 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 217.4 F 0.0
10:00 223.1 F 0.0
15:00 221.1 F 0.0
20:00 218.1 F 0.0
25:00 215.8 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

08/30/2023 01:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____