
Midmark M11 - v1.0.0
Total Cycles: 4799

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2024 04:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 158.0 F 1.6
4:00 184.8 F 5.7
6:00 206.1 F 11.0
8:00 217.2 F 4.9
10:00 231.0 F 9.8
12:00 234.4 F 8.7
14:00 242.9 F 12.2
16:00 240.8 F 10.6
18:00 254.3 F 17.4
20:00 266.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.3 F 28.3
2:00 272.7 F 28.8
3:00 271.5 F 28.0
4:00 272.5 F 28.6

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 202.5 F 0.0
10:00 202.1 F 0.0
15:00 202.9 F 0.0
20:00 208.2 F 0.0
25:00 233.4 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

08/08/2024 05:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____