
Midmark M11 - v1.0.0
Total Cycles: 4539

Name: RITTER 1
Sterilizer ID: V2178198
05/09/2024 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.1 F 0.0
2:00 121.4 F 1.4
4:00 138.4 F 4.2
6:00 166.5 F 8.1
8:00 189.8 F 12.6
10:00 217.6 F 5.8
12:00 223.5 F 10.7
14:00 225.4 F 6.2
16:00 236.8 F 12.0
18:00 236.5 F 9.4
20:00 249.3 F 15.6
22:00 260.9 F 22.0
24:00 271.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 273.4 F 29.8
2:00 271.6 F 29.2
3:00 273.9 F 29.9
4:00 271.8 F 29.2

Min 271.5 F 29.1
Max 273.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.6 F 0.0
5:00 234.0 F 0.0
10:00 220.4 F 0.0
15:00 231.6 F 0.0
20:00 217.9 F 0.0
25:00 225.5 F 0.0
30:00 206.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:43
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:03:00
CYCLE COMPLETE

05/09/2024 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____