
Midmark M11 - v1.0.5
Total Cycles: 3224

Name: RITTER2
Sterilizer ID: V2313228
12/15/2023 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.7 F 0.0
2:00 125.4 F 0.7
4:00 149.2 F 3.0
6:00 173.7 F 6.4
8:00 194.4 F 10.3
10:00 216.7 F 6.8
12:00 226.8 F 8.3
14:00 239.0 F 12.8
16:00 236.4 F 9.5
18:00 238.9 F 10.1
20:00 245.5 F 13.2
22:00 256.6 F 19.2
24:00 266.9 F 25.6
26:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.4 F 29.4
2:00 272.1 F 29.6
3:00 272.2 F 29.8
4:00 271.2 F 29.4

Min 271.1 F 29.1
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 233.9 F 0.0
10:00 208.9 F 0.0
15:00 181.5 F 0.0
20:00 172.9 F 0.0
25:00 171.4 F 0.0
30:00 170.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 26:13
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:04:45
CYCLE COMPLETE

12/15/2023 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____