

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000650  
11:02 19/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:35  
PRESSURIZING 3:36  
STERILIZING 4:40  
136.2 C 222kPa 4:40  
136.3 C 224kPa 5:10  
136.4 C 225kPa 5:40  
136.2 C 222kPa 6:10  
136.3 C 225kPa 6:40  
136.3 C 223kPa 7:10  
136.2 C 221kPa 7:40  
136.3 C 223kPa 8:10  
136.3 C 223kPa 8:10  
VENTING 8:10  
AIR DRYING 9:05  
CYCLE COMPLETE 12:49

-----  
Digital Signature #  
A7AE52681E124861  
-----