
Midmark M11 - v1.0.5
Total Cycles: 1710

Name: RITTER2
Sterilizer ID: V2313228
11/10/2022 03:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 190.3 F 2.9
4:00 205.3 F 8.5
6:00 220.6 F 9.8
8:00 232.8 F 10.2
10:00 233.7 F 8.2
12:00 240.4 F 10.6
14:00 247.9 F 14.7
16:00 260.8 F 21.9
18:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.3 F 29.8
2:00 271.8 F 29.0
3:00 272.1 F 29.7
4:00 272.2 F 29.3

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 232.7 F 0.0
10:00 237.3 F 0.0
15:00 231.0 F 0.0
20:00 227.3 F 0.0
25:00 226.9 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 18:23
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:40
CYCLE COMPLETE

11/10/2022 04:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____