

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.2uS / 0.7ppm  
CYCLE NUMBER 001523  
13:07 10/21/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:59  
PRESSURIZING 6:36  
STERILIZING 8:00  
136.0 C 220kPa 8:00  
135.9 C 218kPa 8:31  
136.2 C 221kPa 9:01  
136.4 C 223kPa 9:31  
136.4 C 223kPa 10:01  
136.4 C 221kPa 10:31  
136.5 C 223kPa 11:01  
136.4 C 221kPa 11:31  
VENTING 11:31  
AIR DRYING 12:39  
CYCLE COMPLETE 41:06

-----  
Digital Signature #  
F451D5641508A4BB  
-----