
Midmark M11 - v1.0.5
Total Cycles: 4109

Name: RITTER2
Sterilizer ID: V2313228
11/12/2024 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.2 F 0.0
2:00 91.4 F 0.0
4:00 111.4 F 0.0
6:00 142.7 F 0.2
8:00 172.3 F 2.2
10:00 190.3 F 5.2
12:00 204.7 F 8.6
14:00 217.5 F 12.4
16:00 223.2 F 6.4
18:00 233.8 F 10.6
20:00 233.6 F 8.4
22:00 243.6 F 12.9
24:00 242.4 F 11.6
26:00 253.9 F 17.6
28:00 264.5 F 23.9
30:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.2
2:00 271.6 F 28.9
3:00 272.4 F 29.0
4:00 271.6 F 29.0

Min 271.4 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 221.2 F 0.0
10:00 197.3 F 0.0
15:00 193.5 F 0.0
20:00 195.8 F 0.0
25:00 203.0 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:03
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:08:33
CYCLE COMPLETE

11/12/2024 01:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____