
Midmark M11 - v1.0.0
Total Cycles: 4144

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2023 05:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.2 F 0.0
2:00 173.4 F 3.9
4:00 203.9 F 8.5
6:00 221.7 F 9.3
8:00 230.5 F 8.4
10:00 240.4 F 12.0
12:00 242.0 F 11.2
14:00 239.4 F 9.8
16:00 253.2 F 16.5
18:00 264.7 F 23.3
20:00 270.8 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.1 F 28.8
2:00 272.4 F 28.8
3:00 272.5 F 28.8
4:00 272.4 F 28.9

Min 271.4 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 256.2 F 0.0
10:00 244.1 F 0.0
15:00 235.9 F 0.0
20:00 228.4 F 0.0
25:00 228.5 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

12/20/2023 06:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____