
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008397
8:37 11/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:06
STERILIZING 3:58
136.0 C 220kPa 3:58
136.4 C 225kPa 4:28
136.4 C 224kPa 4:58
136.0 C 217kPa 5:28
136.1 C 220kPa 5:58
136.1 C 218kPa 6:28
136.4 C 223kPa 6:58
136.3 C 224kPa 7:28
136.3 C 224kPa 7:28
VENTING 7:28
AIR DRYING 8:16
CYCLE COMPLETE 38:17

Digital Signature #
9A645F6E03512409
