
Midmark M11 - v1.0.0
Total Cycles: 1706

Name: RITTER 1
Sterilizer ID: V2178198
03/10/2022 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.4 F 0.0
2:00 204.5 F 3.4
4:00 206.0 F 7.7
6:00 225.5 F 13.2
8:00 230.4 F 8.5
10:00 241.3 F 11.6
12:00 241.6 F 11.5
14:00 239.3 F 9.8
16:00 253.0 F 16.6
18:00 264.6 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.1 F 28.8
2:00 271.3 F 28.0
3:00 271.9 F 28.9
4:00 271.9 F 28.5

Min 271.0 F 28.0
Max 272.1 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 227.8 F 0.0
10:00 224.8 F 0.0
15:00 229.7 F 0.0
20:00 225.0 F 0.0
25:00 226.9 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 2:07
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

03/10/2022 11:48 AM
ACCEPTED: X
REJECTED:
Operator: U02