
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000043
13:07 15/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:05
PRESSURIZING 4:05
STERILIZING 5:24
136.4 C 222kPa 5:24
136.4 C 226kPa 5:54
136.4 C 223kPa 6:24
136.4 C 225kPa 6:54
136.4 C 225kPa 7:24
136.3 C 224kPa 7:54
136.3 C 223kPa 8:24
136.4 C 224kPa 8:54
136.4 C 224kPa 8:54
VENTING 8:54
AIR DRYING 10:01
CYCLE COMPLETE 10:04

Digital Signature #
15CB585094E60E32
