
Midmark M11 - v1.0.0
Total Cycles: 1144

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2021 04:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 168.6 F 2.3
4:00 184.8 F 6.4
6:00 207.2 F 11.9
8:00 220.4 F 6.4
10:00 236.7 F 11.7
12:00 234.4 F 9.1
14:00 239.5 F 9.5
16:00 247.7 F 14.5
18:00 260.3 F 21.4
20:00 271.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.8 F 29.2
2:00 271.8 F 28.7
3:00 272.8 F 29.3
4:00 272.0 F 28.6

Min 271.8 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 226.8 F 0.0
10:00 215.8 F 0.0
15:00 219.1 F 0.0
20:00 218.1 F 0.0
25:00 220.4 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 3:13
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

09/02/2021 05:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____