
Midmark M11 - v1.0.0
Total Cycles: 3267

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2023 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.1 F 0.0
2:00 197.2 F 0.8
4:00 213.1 F 6.3
6:00 223.2 F 11.7
8:00 228.8 F 7.7
10:00 242.1 F 13.4
12:00 242.1 F 11.7
14:00 240.7 F 10.6
16:00 254.9 F 17.8
18:00 266.6 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.1 F 28.8
2:00 272.4 F 28.6
3:00 271.5 F 28.1
4:00 272.3 F 29.0

Min 271.4 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 231.7 F 0.0
10:00 239.4 F 0.0
15:00 227.3 F 0.0
20:00 224.6 F 0.0
25:00 223.5 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

05/23/2023 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____