
Midmark M11 - v1.0.0
Total Cycles: 3263

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2023 11:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.2 F 0.0
2:00 202.1 F 2.1
4:00 213.7 F 8.0
6:00 228.2 F 11.3
8:00 235.1 F 9.9
10:00 235.9 F 8.4
12:00 242.9 F 11.2
14:00 247.7 F 13.8
16:00 260.6 F 20.9
18:00 272.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.7 F 28.6
2:00 271.5 F 27.9
3:00 272.8 F 28.8
4:00 272.0 F 28.1

Min 271.5 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 214.8 F 0.0
10:00 230.1 F 0.0
15:00 227.6 F 0.0
20:00 225.4 F 0.0
25:00 224.2 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 2:44
HEATING: 18:30
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

05/19/2023 12:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____