
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000834
13:39 31/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:36
PRESSURIZING 6:22
STERILIZING 8:02
136.4 C 223kPa 8:02
136.4 C 223kPa 8:32
136.4 C 222kPa 9:02
136.3 C 222kPa 9:32
136.3 C 222kPa 10:02
136.3 C 222kPa 10:32
136.2 C 221kPa 11:02
136.4 C 222kPa 11:32
136.4 C 222kPa 11:32
VENTING 11:32
AIR DRYING 12:34
CYCLE COMPLETE 22:03

Digital Signature #
AD6DC1E89986BBA4
