
Midmark M11 - v1.0.5
Total Cycles: 1670

Name: RITTER2
Sterilizer ID: V2313228
11/02/2022 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.8 F 0.0
2:00 141.5 F 2.0
4:00 151.9 F 5.2
6:00 177.8 F 9.7
8:00 214.6 F 6.9
10:00 220.2 F 8.8
12:00 233.3 F 12.2
14:00 236.1 F 10.5
16:00 234.9 F 8.2
18:00 248.2 F 15.0
20:00 259.2 F 21.4
22:00 269.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.8
1:00 272.0 F 29.9
2:00 272.0 F 29.9
3:00 271.9 F 29.8
4:00 272.1 F 29.7

Min 271.8 F 29.7
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 228.2 F 0.0
10:00 238.4 F 0.0
15:00 227.6 F 0.0
20:00 224.4 F 0.0
25:00 223.1 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:34
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:53
CYCLE COMPLETE

11/02/2022 02:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____