

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1956

Name: RITTER 1  
Sterilizer ID: V2178198  
05/26/2022 09:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.7 F 0.0  
2:00 173.4 F 2.7  
4:00 204.4 F 6.9  
6:00 214.5 F 12.8  
8:00 225.5 F 6.6  
10:00 237.0 F 11.7  
12:00 234.9 F 8.3  
14:00 243.2 F 11.3  
16:00 245.1 F 12.7  
18:00 257.4 F 19.3  
20:00 268.3 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.3 F 29.1  
2:00 271.8 F 28.4  
3:00 272.1 F 28.9  
4:00 272.2 F 28.7  
  
Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 219.8 F 0.0  
10:00 218.6 F 0.0  
15:00 224.8 F 0.0  
20:00 223.5 F 0.0  
25:00 222.2 F 0.0  
30:00 218.9 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 21:06  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:03  
CYCLE COMPLETE

05/26/2022 10:51 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02