
Midmark M11 - v1.0.5
Total Cycles: 4316

Name: RITTER2
Sterilizer ID: V2313228
01/30/2025 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 71.8 F 0.0
4:00 81.6 F 0.9

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 89F
CHAMBER PRESS 1.4PSI
** ITEMS NOT STERILE **

01/30/2025 09:54 AM