

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000669  
10:28 14/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:09  
STERILIZING 4:14  
136.4 C 231kPa 4:14  
136.4 C 233kPa 4:45  
136.5 C 233kPa 5:15  
136.4 C 233kPa 5:45  
136.4 C 233kPa 6:15  
136.4 C 233kPa 6:45  
136.4 C 232kPa 7:15  
136.3 C 232kPa 7:45  
136.3 C 232kPa 7:45  
VENTING 7:45  
AIR DRYING 8:44  
CYCLE COMPLETE 8:45

-----  
Digital Signature #  
D55B486CFEE81100  
-----